



# KERALA GAZETTE

## കേരള ഗസറ്റ്

PUBLISHED BY AUTHORITY

ആധികാരികമായി പ്രസിദ്ധപ്പെടുത്തുന്നത്

Vol. II	}	THIRUVANANTHAPURAM, TUESDAY	20th August 2013	No.	}	34
വാല്യം 2			2013 ആഗസ്റ്റ് 20			
		തിരുവനന്തപുരം, ചൊവ്വ	29th Sravana 1935			
			1935 ശ്രാവണം 29			

## PART III

### Stores Purchase

#### Health Services Department

##### SHORT TENDER NOTICE

No. A7/6548/2012.

25th July 2013.

Competitive sealed tenders are invited for the supply of the following item to the Trauma Care Unit at District Hospital, Palakkad.

Description of the Items—C-arm Image Intensifier.

PAC—₹ 10 Lakhs.

EMD—₹ 10,000.

Cost of tender form:

Original—₹ 1,500+VAT @ 12.5%.

Duplicate—₹ 800+VAT @ 12.5%.

The tender should be superscribed “Tender for the supply of C-arm Image Intensifier” and should be addressed to the Superintendent, District Hospital, Palakkad. The tender forms can be purchased from the Office of the Superintendent, District Hospital from 10 a. m. to 3 p. m. on any working day till 29-8-2013. The last date and time for receipt tenders will be 29-8-2013, 2 p. m. The tenders received within the stipulated time will be opened at 3.30 p. m. on the same day in the presence of such tenderers or their authorized representatives who may be present at that time.

The tender forms are not transferable. Cheque, postage stamps etc. will not be accepted towards the cost of the forms nor will the form be sent through VPP or by post. The cost of the tender forms will not be refunded.

### **Special Conditions**

The intending tenderers along with the tender should submit D. D. drawn in favour of Superintendent, District Hospital, Palakkad as EMD.

The successful tenders should execute an agreement as in appendix XIV of store purchase manual and will remit security deposit of 5 % of the total amount by way of D. D. drawn in favour of Superintendent, District Hospital, Palakkad.

The items should be in conformity with the specification attached. The rate should be quoted inclusive of all taxes levies, exice duty, entry tax etc. and the items should be supplied at door District Hospital, Palakkad.

Superintendent, District Hospital, Palakkad, reserves the right to accept, reject or to cancel the tender received without assigning any reason thereof. All other general conditions on tender will also be applicable.

The rate quoted shall be with comprehensive warranty for a period of 3 years. CMC and AMC rate for further a period of 7 years shall also be specified.

### **SPECIFICATION**

#### **Image Intensifier-C-Arm with Monitor Specification**

- C-ARM Mobile Intensifier with rotating anode X-ray Tube.
- 9 inch Image Intensifier.
- High Resolution CCD Camera.
- Image orientation left/right, up/down and rotation.
- Double monitors, Flicker free 17" at 75Hz. Or above with flat screen.
- Digital display of KVP, FMA, MAS and flouro timer.
- Tube temperature sensor with Thermal Safety cut off.
- X-ray Generator should be high frequency 40 Khz system.
- Automatic brightness stabilizer (ABS).
- Motorised up/down movement.
- Fluroscopy and radiography mode.
- Double focal spot X-ray tube.
- Voltage stabilizer 2 KVA-1 No.
- Light weight led Aprons-4 Nos.
- Manufacturer should have AERD Certifications.
- Should have service entire in Kerala.

District Hospital,  
Palakkad.

(Sd.)  
*Superintendent.*